

An introduction to inverse treatment planning / dose optimization (for photons and ions)

NIKLAS WAHL



This project has received funding from the European Union's Horizon 2020 research and innovation programme under grant agreement No 101008548

May 18, 2021

Heavy Ion Therapy
Masterclass school

An introduction to inverse treatment planning / dose optimization (for photons and ions)

Niklas Wahl

Group Radiotherapy Optimization
Department of Medical Physics in Radiation Oncology

DKFZ Heidelberg

HIRO

Heidelberg Institute
for Radiation Oncology

National Center for
Radiation Research in
Oncology Heidelberg

supported by:
German Cancer Research Center (DKFZ)
Heidelberg University Hospital
Heidelberg Ion Therapy Center (HIT)
Medical Faculty Heidelberg


dkfz.

GERMAN
CANCER RESEARCH CENTER
IN THE HELMHOLTZ ASSOCIATION



Research for a Life without Cancer

Radiation therapy




Adapted from W. Schlegel & A. Mahr: 3D Conformal Radiation Therapy Springer Multimedia DVD

A treatment plan should ...

- ... fulfil clinical requirements (“The physician prescribes”) ...
- ... based on biological processes (cell death) ...
- ... induced by chemical & physical processes (reactions & interactions) ...
- ... by means of numerical simulation (dose calculation / optimization)

→ Interdisciplinary **inverse problem**

Imaging: X-ray




[1] anonym (<https://commons.wikimedia.org/wiki/File:Roentgen2.jpg>), „Roentgen2“, marked as public domain, more details on Wikimedia Commons: <https://commons.wikimedia.org/wiki/Template:PD-EU-no author disclosure>

[2] Wilhelm Röntgen; current version created by Old Moonraker. (https://commons.wikimedia.org/wiki/File:X-ray_by_Wilhelm_Röntgen_of_Albert_von_Kölliker's_hand_-_18960123-02.jpg), „X-ray by Wilhelm Röntgen of Albert von Kölliker's hand - 18960123-02“, marked as public domain, more details on Wikimedia Commons: <https://commons.wikimedia.org/wiki/Template:PD-old>

[3] Photograph: JonathunderMedal: Erik Lindberg (1873-1966) (https://en.wikipedia.org/wiki/File:Nobel_Prize.png), „Nobel Prize“, marked as public domain, more details on Wikimedia Commons: <https://commons.wikimedia.org/wiki/Template:PD-US>

Imaging: Computed tomography


“3D X-ray”




[1] FDA – Radiation emitting products – Medical X-ray Imaging – What is Computed Tomography? - Accessed from <https://www.fda.gov/radiation-emitting-products/medical-x-ray-imaging/what-computed-tomography> on 15.02.2021.

[2] daveynin from United States (https://commons.wikimedia.org/wiki/File:UPMCEast_CTscan.jpg), „UPMCEast CTscan“, <https://creativecommons.org/licenses/by/2.0/legalcode>


CT scans for treatment planning



Segmentation: Delineating volumes of interest




The ideal dose distribution




- High / prescribed dose in the tumor
- No dose in normal tissue

Modern 3D-planning with photons

Photon beam




Collimator



→ Adaptation of the photon beam to the tumor shape

A realistic dose distribution




Can we do better?

The concept of the „beamlet“



- „Multi-leaf“ collimator are able to generate fine beams (let's call them beamlets)
- We calculate their dose for unit intensity using various algorithms, e.g.:
 - **Analytical Pencil beam**
Precomputed / measured dose curves in water are “scaled” to the patient
→ deterministic, very quick, but inaccurate
 - **Monte Carlo**
Simulation of individual particle trajectories (“histories”) through the patient
→ stochastic, slow, but mostly more accurate

→ we are able to simulate and “modulate” our beams

#3



Intensity-modulation with pencil-beams




Different beamlet **weights w**
= Intensity-modulated field

- Number of beamlets: ~100-1000 per field
- Number of fields: 5 to 12

→ We can't do it by hand

The ideal dose distribution (again)



- Write ideal dose as a vector:

$$d^* = (\dots, 0, \dots, d^p, \dots)^T \in \mathbb{R}_+^I$$

I = number of voxels

- Put the beamlets in a *dose influence matrix*:

$$D \in \mathbb{R}_+^{J \times I}$$

J = number of beamlets

- Fluence weight vector:

$$w \in \mathbb{R}_+^J$$

→ We would like to solve $d^* = Dw$ for w

Finding the right vector w^*

- $d^* = Dw$ has no solution: $d^* = Dw + \epsilon$
- Approximate $d^* \rightarrow$ We need to minimize ϵ !
- An exemplary straightforward optimization approach:
weighted/penalized least squares

$$w^* = \arg \min_w (Dw - d^*)^T P (Dw - d^*)$$


$$\text{s. t. } w \geq 0$$

$$P = \text{diag}(p_1, p_2, \dots, p_I)$$


Optimize for w^*

dimensionality = number of pencil beams ($\sim 10^3$ - 10^4) \rightarrow Quasi-Newton Method

1. Initial solution
2. Obj. func. calc.
3. Grad. calc.
4. Search direction
5. Step length
6. Go to step #2




Intensity modulated treatment plan – dose distribution



Let's move to hadrons & ions...

The “old” way: Passive Scattering




RL Maughan, MJ Hardy, MJ Taylor, J Reay, R Amos: Radiation Shielding and safety for particle therapy facilities, IPEM Report 75 Ed. 2

Translation to intensity-modulated particle therapy (IMPT)

→ We do not need a collimator


- Hadron / ion accelerators (synchrotrons / cyclotrons) produce “natural” beamlets
- improved dose distribution (Bragg-Peak)
- can also be simulated with deterministic or MC algorithms
- additional modulation in depth
→ much higher number of beamlets per beam



Examples: Liver patient


photons

axial plane z = 317.5 [mm]




protons

axial plane z = 317.5 [mm]



carbon ions

axial plane z = 317.5 [mm]




How do we analyze a dose / quantify a plan?

Evaluation of the
2D tomographic images


Dose statistics:
Mean, maximum, minimum dose

Dose-volume histograms
2D display of the 3D dose distribution

Complication / Control Models
(N)TCP, mostly derived empirically



Dose-volume histograms



x% of the volume
obtains at least d%
of the prescribed
dose

TCP / NTCP models

- Tumor control & normal tissue complication often modeled with logistic functions
→ Lyman-Kutcher-Burman – LKB
- Empirically determined, more recently often ML/AI-driven
- May be used directly in planning, but more commonly used to assess plan quality and “design” prescriptions




Image from Tseng H-H, Luo Y, Ten Haken RK and El Naqa I (2018) The Role of Machine Learning in Knowledge-Based Response-Adapted Radiotherapy. *Front. Oncol.* 8:266. doi: 10.3389/fonc.2018.00266

More complex prescriptions...

- The prescription can be much more complex / abstract than the initial example
- Prescriptions need to be translated into the language of the optimizer (math)



„Keep the mean dose to the parotid gland low....“

„while achieving a coverage with 60 Gy in the tumor“

„do not exceed a dose of 10 Gy in the brainstem“

...require more complex optimization problems

„Keep the mean dose to the parotid gland low....“

„while achieving a coverage with 60 Gy in the tumor“

„do not exceed a dose of 10 Gy in the brainstem“

Objective (minimize)

$$f_1 = \frac{1}{N_S} \sum_{i \in S} d_i$$

Objective (minimize)

$$f_2 = \frac{1}{N_S} \sum_{i \in S} (d_i - \hat{d})^2$$

Constraint (enforce)

$$c_1 = d_{max} + \kappa \log \left(\sum_{i \in S} \frac{d_i - d_{max}}{\kappa} \right)$$


- Optimizers:
 - interior-point method
 - Sequential quadratic programming (e.g. in RayStation)

$$\min_{w \in \mathbb{R}^B} f(w) = \sum_n p_n f_n(w)$$

$$\text{s.t.} \quad c_k^l \leq c_k(w) \leq c_k^u$$

$$0 \leq w$$

Example of fulfilled constraints in a DVH



- DVH constraints on cerebellum & parotid glands
- Maximum dose constraint on brainstem

In practice: „Sea“ of objective functions for targets and healthy tissue

- Non-linear constrained optimization problem

$$w^* = \arg \min f(d(w)) = \sum_k^K p_k f_k(d(w))$$

$$\text{s. t. } c_q^L \leq c_q(d(w)) \leq c_q^U$$

$$w \geq 0$$

- Example from matRad paper:


| objectives | | constraints | |
|-------------------------------|---|------------------------|--|
| $f_{sq \text{ deviation}}$ | $= \frac{1}{N_S} \sum_{i \in S} (d_i - \hat{d})^2$ | | |
| $f_{sq \text{ under dosage}}$ | $= \frac{1}{N_S} \sum_{i \in S} \Theta(\hat{d} - d_i)(d_i - \hat{d})^2$ | $c_{min \text{ dose}}$ | $= d_{min} - \kappa \log \left(\sum_{i \in S} e^{\frac{d_{min} - d_i}{\kappa}} \right)$ |
| $f_{sq \text{ over dosage}}$ | $= \frac{1}{N_S} \sum_{i \in S} \Theta(d_i - \hat{d})(d_i - \hat{d})^2$ | $c_{max \text{ dose}}$ | $= d_{max} + \kappa \log \left(\sum_{i \in S} e^{\frac{d_i - d_{max}}{\kappa}} \right)$ |
| f_{mean} | $= \frac{1}{N_S} \sum_{i \in S} d_i$ | c_{mean} | $= \frac{1}{N_S} \sum_{i \in S} d_i$ |
| f_{EUD} | $= \left(\frac{1}{N_S} \sum_{i \in S} d_i^a \right)^{\frac{1}{a}}$ | c_{EUD} | $= \left(\frac{1}{N_S} \sum_{i \in S} d_i^a \right)^{\frac{1}{a}}$ |
| $f_{min \text{ DVH}}$ | $= \frac{1}{N_S} \sum_{i \in S} \Theta(\hat{d} - d_i) \Theta(d_i - \tilde{d})(d_i - \hat{d})^2$ | $c_{min \text{ DVH}}$ | $= \frac{1}{N_S} \sum_{i \in S} \Theta(\hat{d} - d_i)$ |
| $f_{max \text{ DVH}}$ | $= \frac{1}{N_S} \sum_{i \in S} \Theta(d_i - \hat{d}) \Theta(\tilde{d} - d_i)(d_i - \hat{d})^2$ | $c_{max \text{ DVH}}$ | $= \frac{1}{N_S} \sum_{i \in S} \Theta(d_i - \hat{d})$ |

Most important thing to never forget: It's a trade-off

- Trade-off between target coverage and sparing of normal tissues
- Trade-off between sparing of different organs at risk

Multicriteria decision/planning/optimization problem

More options to solve this problem, e.g., with Pareto surface approximation & exploration




Head and neck animation courtesy of Dr. Sebastiaan Breedvelt @ Erasmus MC Cancer Institute Rotterdam

https://sebastiaanbreedveld.nl/rt_tradeoffs.html


**Now we have fully understood and solved the
problem of inverse treatment planning using
intensity modulation for photons & ions?**

Problem: Dealing with uncertainties



Animations courtesy of Paul Merca & Markus Stoll

Margins in treatment planning




- **GTV = Gross tumor volume**
tumor volume that is visible on the images
- **CTV = Clinical target volume**
includes the GTV and regions where invisible tumor tissue is expected
- **PTV = Planning target volume**
safety margin to take uncertainties into account

W. Schlegel & A. Mahr: 3D Conformal Radiation Therapy Springer Multimedia DVD


ICRU report 50

Margins – really?

- The “original” margin recipe for photon therapy:
“**Minimum dose to CTV is 95% for 90% of population**” $2.5 \sigma_{\text{sys}} + 0.7 \sigma_{\text{rand}} - 3\text{mm}$




→ Not applicable for protons / ions!




[3] Steitz et al., *Radiation Oncology* 2016, 11:134

Difference in biological effect

physical dose



Biologically effective dose



→ Physical doses similar to photons induce different biological effect!

Summary

- In treatment planning we try to find the best approximation to an ideal dose distribution / prescription
→ Multiple factors (biology, importance of objectives / organs, etc.)
- The problem is approached by optimization techniques (inverse planning)
- It is a multicriteria problem that let's the planner choose trade-offs
- Ion therapy has its advantages but also pitfalls during planning, most notably
 - Biology
 - Localization (NT sparing & tumor coverage) \Leftrightarrow Uncertainties

THANK YOU

This presentation was prepared for the HITRIplus Heavy Ion Therapy Masterclass School 2021.

Usage in any another context must be confirmed with the authors/presenters beforehand and must always include appropriate attribution.

